JCAAI

Joint Council of Allergy, Asthma and Immunology

50 N. Brockway St.
Suite 3-3
Palatine, IL 60067
847/934-1918
FAX 847/934-1820

October 1, 1999

Dockets Management Branch (HFA-305)
Food and Drug Administration
5630 Fishers Lane, Room 1061
Rockville, MD 20852

RE: Docket No. 98N-03 13
RIN 0910 – AB74
Surgeon's and Patient Examination Gloves; Reclassification
Proposed Rule

To Whom It May Concern:

The Joint Council of Allergy, Asthma and Immunology (JCAAI) appreciates the opportunity to comment on this proposed rule. JCAAI is an organization of allergists and immunologists whose sponsors are the American Academy of Allergy, Asthma and Immunology (AAAAI) and the American College of Allergy, Asthma and Immunology (ACAAI). JCAAI represents over 4,000 physicians, most of whom are board certified by the American Board of Allergy and Immunology, a conjoint board of the American Board of Internal Medicine and the American Board of Pediatrics.

Enclosed are the comments of Dr. Kevin J. Kelly, which we solicited, on this issue. These comments completely reflect the views of the Joint Council of Allergy, Asthma and Immunology. We appreciate your consideration of our comments.

Sincerely,

Daniel Ein MD

Daniel Ein, M.D.
President

DE/sg
enclosure

MEDICAL COLLEGE OF WISCONSIN

Jordan N. Fink, M.D.
Professor & Chief, Allergy-Immunology Division

Kevin J. Kelly, M.D.
Professor & Chief, Pediatric Allergy

Michael C. Zacharisen, M.D.
Assistant Professor of Pediatrics

Asnani M. Chiu, M.D.
Assistant Professor of Medicine

Asthma & Allergy Center
Allergy-Immunology Division
Departments of Medicine and Pediatrics

September 15, 1999

Dockets Management Branch (HFA-305)
Food and Drug Administration
5630 Fishers Lane
Room 1061
Rockville, MD 20852

RE: Docket No. 98N - 03 13
RIN 0910 - AB74
Surgeon's and Patient Examination Gloves; Reclassification
Proposed Rule

To whom it may concern:

I am writing in regards to the proposed regulation to reclassify all surgeon and patient examination gloves to class II devices and set allowable limits of powder and protein in latex surgeon and examination gloves used for patients in the medical field. I would commend the proposed rule change as a "next good step" after the labeling requirement for latex containing products that went into effect September 30, 1998. However, there continues to be a number of problems that exist with the proposed ruling of which further work needs to continue

First, the level of 1200 milligrams of protein in a glove is an unacceptably high level of protein and is only chosen due to the insensitivity of the proposed lab method of protein quantification. In addition, there is no attempt in this process to identify the allergenic protein and distinguish those proteins from non allergenic protein. This requires the USC of specific immunologic assays and not the modified Lowry Test. The modified Lowry Test is insensitive and will allow unacceptably high levels of allergen to persist in latex gloves under this ruling. Since patients have reacted adversely in skin testing to levels of protein as low as one microgram or smaller, the allowance of 1200 milligram is excessive.

Second, the reduction of powder from 260 milligram (on average) to a 120 milligram is an unacceptably high level especially in light of the continued allowance of protein levels in the glove under this proposed rule. Data from Europe suggest that release of allergen into the air at levels of 0.6 nanograms per cubic meter of air is enough to invoke symptomatic disease in health care workers. This is clearly excessive powder to bind allergen and exceed levels that produce symptoms of latex allergy.

9000 West Wisconsin Avenue
P.O. Box 1997
Milwaukee, Wisconsin 53201
(414) 266-6840
FAX: (414) 266-6437

Therefore, the continued work on quantitative immunologic assays and alternatives to powder as a lubricant for the donning of latex gloves is necessary and should be mandatory in the future. Although supportive of the overall attempt to place limits of powder and protein content of latex gloves, this rule must be looked at as merely a first step in reducing allergic reactions in the sensitized health care workers who are exposed to those materials.

Extending the proposed leve of protein to latex materials other than gloves should be considered, especially for those latex products manufactured by a dipping process. In light of the majority of reactions reported to the Food and Drug Administration being related to allergic reactions from dipped latex products, extending this rule to those products makes good sense.

Although the mathematical discussion regarding reduction of risk of allergic reactions is clear, given the liberal new allowable limits of powder and protein in the gloves, the impact of this rule is likely to be overestimated. Since latex allergic patients adversely react below levels produced by the new standard, it is highly unlikely that the number of allergic reactions will be reduced until more drastic measures are imposed. The reduction in allergic reactions is unlikely to occur until a critical reduction (nondetection) in the level of allergen (not just protein) is achieved.

If my statements are not clear or I can be of further help in clarifying the issues, please feel free to contact me at (414) 266-6840.

Sincerely,

Kevin J. Kelly, M.D.
Professor of Pediatrics & Medicine
Director, Pediatric Allergy & Immunology
Medical College of Wisconsin

Attach the Airborne Express Shippers Label within the dotted lines.



United States Shipping

1. Complete applicable white sections of the U.S. Airbill. Sign and date the Airbill at the Sender's Signature line. Please press hard.
2. Peel off protective covering from back of Airbill.
3. Affix Airbill to envelope within dotted lines shown.
4. **When using a Drop Box – follow special instructions on the Drop Box.**

International Shipping

Includes Canada & Puerto Rico

Must be typed

1. Complete applicable white sections of the International Express Airbill. Sign and date the Airbill at the Sender's Signature line.
2. Place Airbill in plastic sleeve.
3. Peel off bottom portion from back of plastic sleeve. **Do not seal top portion of the plastic sleeve to the envelope.**
4. Affix bottom portion to envelope within the dotted lines shown. **Airborne driver must sign Airbill before sealing.**

Limitation on Contents

The maximum acceptable contents of a Letter Express is forty (40) 8-1/2 x 11 pages. If the gross weight of the contents, envelope and airbill exceeds 1/2 pound, the next higher rate will apply. Contents must be of a size and shape which fit the envelope and allow it to be securely sealed without damage. Cash or cash equivalent should not be tipped. Items of high intrinsic value should not be shipped in Letter Express packaging.

Limitations of Liability

Liability of Airborne Express is limited on Letter Express to $100.00 U.S.D., unless a higher value is declared for carriage on our airbill. The maximum declared value on the Letter Express is $500.00 U.S.D. Airborne Express shall not be liable in any event for special, incidental or consequential damages, including but not limited to loss of profits or income. Services are provided as defined in the current Airborne Express Service Guide (subject to change without notice). Copies are available upon request.

To reach your local CUSTOMER SERVICE CENTER **Call 1-800-AIRBORNE (1-800-247-2676) U.S. only.**